

Data Sheet:

Chair-Check® II Model 72010

Rugged, Versatile Fall Monitoring

The Chair-Check II control unit is used in conjunction with a Bed-Check® Sensormat® pad as a system to effectively monitor patients at risk of falling or residents in a bed, chair, wheelchair, or on a toilet. Chair-Check II provides an audible alarm to alert the caregiver when the person attempts to get up unassisted.

Product Description

- Monitor patient or resident in chair, toilet, wheelchair, or bed, using any Bed-Check® Sensormat® pad
- Control unit uses a 9V alkaline battery for a long usage life
- Secured battery containment discourages tampering
- Non-latching alarm allows monitoring to restart automatically when individual's weight is back on the Sensormat pad, and does not require caregiver intervention
- Configure Chair-Check II settings to the individual's needs with 3 selectable alarm tones, 2 volume settings, and a 0–5 second delay
- Assembled in the USA; Two-year limited warranty

Features and Operation

The Chair-Check II is the most versatile control unit in Bed-Check's product line.

- Monitoring begins automatically when the patient's weight is placed on the Sensormat pad
- Caregiver can put the system on Hold while repositioning the patient
- Controls for hold, tone, and delay may be easily changed by the caregiver
- Multiple tones allow the unit to be distinguished from other alarms or other Chair-Check II monitors
- Low battery and charge alert. Secured battery access
- 9V battery or AC operation allows for use in portable or fixed applications
- Chair Check II features configurable delay settings
 - Delay is 0 seconds for most seated applications, since the patient is postured for quick stand up.
 - Delay is 1, 3, or 5 seconds may be set for other circumstances, such as bed monitoring
- Utilizes pressure-sensitive Sensormat pads available for chair, bed or toilet monitoring. Not intended for use with other manufacturers' pads
- Non-latching alarm action. The alarm ceases and monitoring automatically resumes when weight is again sensed, reducing caregiver intervention compared to string-activated alarms
- Conforms to UL® 1069 and CSA C22.2 No. 205R Standards

Accessories

- Holsters: Wire basket holster for easier mounting on chairs and headboards, with a 1" or 2" depth hanger.

